WATER QUALITY / DATA SHEET

SS316-01-WIFI-4

Ammonia Nitrogen Transmitter



Online ammonia nitrogen and temperature monitoring with automatic temperature compensation, manual pH compensation and WiFi upload to a local platform, cloud platform or customer server.

Overview

- Three ranges: 0–10, 0–100 or 0–1000 mg/L.
- Reagent-free monitoring with manual pH compensation.
- WiFi upload; 2 s acquisition; 10 s default upload interval.

Maximum Range	Resolution	Power Supply	Upload
0–1000 mg/L	0.01 / 0.1 mg/L	DC 10–30 V	WiFi

Key Advantages

- Automatic temperature and manual pH compensation.
- Reagent-free online concentration monitoring.
- Three selectable measurement ranges.
- 802.11b/g/n; LAN and WAN communication; DNS support.
- IP68 probe and IP65 conversion module.

Application Scenarios

- Online water-quality monitoring.
- Continuous ammonia nitrogen measurement.
- Immersion or pipeline installation.
- Use in compatible freshwater matrices.

SYSTEM DATA

Technical Specifications

Item	Specification
Model	SS316-01-WIFI-4
Measured parameter	Ammonia nitrogen: nitrogen present as NH3 and NH4+
Range / resolution	0–10 mg/L: 0.01 mg/L; 0–100 mg/L: 0.01 mg/L; 0–1000 mg/L: 0.1 mg/L
Measurement error	±3% FS at 25 °C in laboratory conditions
Response time	<30 s at 25 °C in laboratory conditions
Repeatability error	≤3% at 25 °C in laboratory conditions
Supply voltage	DC 10–30 V
Power consumption	0.5 W
Data upload	WiFi 802.11b/g/n; 2 s acquisition; upload 1–65535 s, default 10 s
Probe operating temperature	0–50 °C
Housing environment	−40 to +60 °C; 0–95% RH, non-condensing
Ingress protection	Probe: IP68; transmitter housing: IP65
Probe cable	5 m standard; 10 / 15 / 20 m customizable
Probe pressure rating	0.2 MPa

Item	Specification
Housing dimensions	117 × 87 × 43 mm maximum
Probe dimensions / thread	266 mm length; 212 mm body; Ø30 mm; NPT 3/4 at both ends
Temperature range / error	−20 to +80 °C; ±0.3 °C (probe operating limit: 0–50 °C)

MEASUREMENT DATA


Electrode Service and Monitoring

Electrode maintenance	Requirement
Typical service life	Ammonia nitrogen electrode: 3–6 months. Replace aged electrodes promptly.
Calibration interval	Calibrate before each use; for continuous immersion, calibrate every 2–3 weeks and adjust for fouling or chemical deposits.

Ordering model: SS316-01-WIFI-4. Specify the 0–10, 0–100 or 0–1000 mg/L ammonia nitrogen range and cable length. The system includes the probe and wall-mounted WiFi transmitter.

MECHANICAL DATA

Dimensions and Installation



Dimensions in millimetres: probe length 266; body length 212; diameter Ø30; wrench flats 28; NPT 3/4 threads at both ends. Module housing: 117 × 87 × 43 maximum. Wall mounting spacing: 105; use Ø5 holes with the supplied plugs and screws.

For immersion installation, route the probe cable through a waterproof support bracket and seal the top 3/4-inch threaded connection with PTFE tape. For pipeline installation, connect via the probe's 3/4-inch thread. Keep cable connectors out of water until fully locked.

ELECTRICAL CONNECTION

Power Connection and Start Up

Connection	Instruction
Power	DC 10–30 V; supplied adapter: 12 V
Start up	Connect the adapter to the device, then energize
Data availability	Wait 1–3 minutes, then view readings on the platform or receiving server

COMMUNICATION DATA

WiFi Configuration and Data Upload

Power the device using the supplied adapter. Use the supplied Android configuration app with Bluetooth enabled. Scan for WIFIOPEN followed by the device address; connect with default password 12345678. Read parameters, edit settings and send them to the device. Read back to verify changes. WiFi supports WEP / WPA-PSK / WPA2-PSK security and WEP / TKIP / AES encryption.

Set the destination IP address or domain and matching listening port: local monitoring software defaults to 2404; cloud platform defaults to 8020. For a static IP, enter the assigned IP address, subnet mask and gateway, then select static IP mode. Upload interval: 1–65535 s, default 10 s. Operation password: 1–8 numeric digits, default 12345678. Customer platforms can use the supplied secondary-development documentation.

Local software runs on a customer PC or server; cloud monitoring requires no customer server. Both support web monitoring, limit alarms, email alerts, history export and permission management. Local SMS alerts require an additional SMS modem.

CALIBRATION & MAINTENANCE

Probe Cleaning and Calibration

Remove the transparent electrode protective sleeve before measurement. Check the sensing end for bubbles; if present, shake the device downward to remove them. Rinse the probe head with clean water after use. Protect the membrane from sharp objects and store unused electrodes in diluted standard solution.

After prolonged storage, activate first in 10 mg/L standard for at least 12 hours, then in 1000 mg/L standard for 1–2 hours. Before testing, immerse the sensing end in deionized water for 5 minutes while stirring; repeat with fresh water for thorough rinsing. After a short idle period, soak in deionized water before measurement or calibration.

Do not use this device as a safety device or emergency stop device, or in applications where equipment failure could cause personal injury. Read the operating instructions before installation, use or servicing.

Avoid strong acids/alkalis, strong oxidizing or reducing media, hydrofluoric acid, sulfides, bromides, iodides and organic solvents. Complex sample chemistry can damage the ion membrane. Potassium ions cause substantial interference; seawater and high-potassium samples are not recommended. Package: transmitter, WiFi transmitter module, 12 V power adapter, 5 m cable, two plugs, two screws, nylon protective net, certificate and warranty card. Warranty: 24 months for verified defects under normal use; the ammonia nitrogen electrode is a consumable and excluded.

CONTACT / SUPPORT

SpeSense Technologies Inc.
Add: No. 9 Yuhe Road, Guitang Subdistrict, Yuhua District, Changsha City
Tel: +86-15386478166
Email: info@spesense.com
Web: www.spesense.com

WHATSAPP WECHAT

WhatsApp / WeChat